

RO•ILS[®]

RADIATION ONCOLOGY
INCIDENT LEARNING SYSTEM

Sponsored by ASTRO and AAPM

CLARITY

PSO

A Patient Safety Organization

AGGREGATE DATA REPORT

Quarter 3, 2022

July 1 - September 30, 2022

Patient Safety Work Product


CLARITY PSO, a Division of Clarity Group, Inc.
8601 W Bryn Mawr Ave • Suite 110 • Chicago, IL 60631
T: 773.864.8280 • F: 773.864.8281 • www.claritypso.com

Aggregate Report Card

The timeframes utilized throughout this report are based on the timepoint of event submission (i.e., when the event was initially entered into the RO-ILS Portal, when the “Submit Event” form was completed). The national aggregate database only includes events that have been reviewed and reported to the PSO, which occurs after submission. All information in this report represent data in the national database. For this reason, the amounts provided for previous timeframes may change as older events are reviewed and reported to the PSO.


METRIC	Q2 2022	Q3 2022	AGGREGATE HISTORICAL SUM
Total Number Of Events	1646	1115	25883
Therapeutic Radiation Incident	189	184	3364
Other Safety Incident	245	135	3200
Near Miss	206	124	4701
Unsafe Condition	437	290	4608
Operational/Process Improvement	569	382	10010
Most Commonly Identified Workflow Step Where Event Occurred	Treatment Delivery Including Imaging 33% (539/1646)	Treatment Delivery Including Imaging 35% (388/1115)	Treatment Delivery Including Imaging 30% (7686/25883)
Most Commonly Identified Workflow Step Where Event was Discovered	On-Treatment QA 26% (434/1646)	Treatment Delivery Including Imaging 26% (293/1115)	Treatment Delivery Including Imaging 29% (7584/25883)
Most Commonly Identified Treatment Technique	IMRT/VMAT 43% (708/1646)	IMRT/VMAT 43% (483/1115)	IMRT/VMAT 34% (8737/25883)
Most Commonly Identified Dose Deviation for Therapeutic Radiation Incidents that Did Not Effect Multiple Patients	≤ 5% maximum dose deviation to target 74% (50/68)	≤ 5% maximum dose deviation to target 44% (34/77)	≤ 5% maximum dose deviation to target 82% (1325/1612)

Aggregate: Total Number of Events Reported to PSO by Quarter Submitted



This graph displays how many events were submitted, reviewed, and reported to the PSO per quarter based on the date of submission.


Aggregate: Event Classification



This graph depicts the distribution of data for data element #104 (*Event_Classification*) on a bi-annual basis. The "Operational/Process Improvement" answer option became available in the Q3 2016 data element update.


■ Therapeutic Radiation Incident
 ■ Other Safety Incident
 ■ Near-miss
 ■ Unsafe condition
 ■ Operational/Process Improvement

Aggregate: Workflow Step(s) Where Event Occurred



This graph depicts the frequency of answer options for data element #208 (*Occurred_Workflow*) based on the calendar year it was submitted. Users have the option to select more than one answer option, therefore, there is a potential for each year's total percentage to be greater than 100%. As a part of the Q3 2016 data element update, this question became required and an answer option of "Outside the Radiation therapy Workflow or Other" was added.

Aggregate: Workflow Step Where Event Discovered




This data is based on data element #207 (*Discovered_Workflow*), presented in yearly increments. As a part of the Q3 2016 data element update, this question became required and an answer option of "Outside the Radiation therapy Workflow or Other" was added.

Aggregate: Workflow Step Where Event Occurred & Event Discovered

<u>Quarter</u>	<u>Discovered Workflow</u>	<u>Occurred Workflow</u>									Grand Total
		Before Simulation	Pre-planning Imaging and Simulation	Treatment Planning	Pre-Treatment QA Review	Treatment Delivery Including Imaging	On-Treatment QA	After Treatment Course is Finished	Equipment and Software QA	Outside The Radiation Therapy Workflow Or Other	
2022 Q3	Before Simulation	95%	5%								100%
	Pre-planning Imaging and Simulation	57%	39%	3%	1%						100%
	Treatment Planning	6%	39%	45%	7%				1%	2%	100%
	Pre-Treatment QA Review	5%	7%	47%	34%	2%	4%			2%	100%
	Treatment Delivery Including Imaging	2%	5%	21%	11%	59%	2%			2%	100%
	On-Treatment QA	0%	1%	3%	1%	60%	34%		0%	1%	100%
	After Treatment Course is Finished		5%	5%	14%	14%	18%	45%			100%
	Equipment and Software QA					33%			67%		100%
	Outside the Radiation Therapy Workflow or Other	1%		2%	1%	7%	1%	1%	1%	87%	100%


This graph displays the overlap of where in the workflow events occurred (208. *Occurred_Workflow*) and where it was discovered (207. *Discovered_Workflow*) for the most recent quarter. The percentage is based on the total number of events that were discovered in that workflow step (e.g., of events that were discovered “before simulation”, X% occurred “before simulation”).

Aggregate: Treatment Technique(s)



This graph demonstrates the distribution of answers for data element #106 (*Tx_Technique*) per quarter for the last two quarters. Users may be select more than one technique if pertinent to the event and thus the quarterly percentage total may be greater than 100%.

Aggregate: Significance Scale



Quarter	Mild	Moderate	Severe
2018 Q1	82%	16%	2%
2018 Q2	85%	13%	2%
2018 Q3	84%	14%	3%
2018 Q4	83%	15%	2%
2019 Q1	83%	15%	2%
2019 Q2	81%	16%	2%
2019 Q3	86%	12%	2%
2019 Q4	81%	17%	3%
2020 Q1	79%	17%	4%
2020 Q2	76%	20%	4%
2020 Q3	78%	20%	3%
2020 Q4	81%	16%	2%
2021 Q1	79%	18%	3%
2021 Q2	82%	15%	3%
2021 Q3	80%	17%	3%
2021 Q4	80%	18%	2%
2022 Q1	83%	15%	2%
2022 Q2	83%	15%	2%
2022 Q3	86%	12%	2%

This graph shows the frequency of answers for data element #225 (*Significance_Scale*) which asks, “In terms of risk to patient safety, how significant was this event?” This question was introduced in Q3 2016. The graph and table provide the severity percentages per quarter since 2017.